

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066635.D
 Acq On : 16 Apr 2021 18:54
 Operator : JC/MD
 Sample : PB135768ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135768ZHE#05

Quant Time: Apr 17 05:19:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	247251	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	394584	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	380194	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	157968	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	185034	49.61	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.22%
35) Dibromofluoromethane	7.516	113	145201	53.06	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.12%
50) Toluene-d8	10.029	98	578683	53.96	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.92%
62) 4-Bromofluorobenzene	12.346	95	190160	50.43	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.86%
Target Compounds						
					Qvalue	
16) Acetone	3.756	43	25793	11.23	ug/l	99
18) Methyl Acetate	4.263	43	4718	1.17	ug/l #	60
20) Methylene Chloride	4.475	84	14447	4.10	ug/l	91
43) Isopropyl Acetate	8.095	43	26686	3.54	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066635.D
Acq On : 16 Apr 2021 18:54
Operator : JC/MD
Sample : PB135768ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135768ZHE#05

Quant Time: Apr 17 05:19:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

