

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066636.D
 Acq On : 16 Apr 2021 19:18
 Operator : JC/MD
 Sample : PB135768ZHE#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135768ZHE#06

Manual Integrations
 APPROVED

MMDadoda
 4/19/2021 2:59:27 PM

Quant Time: Apr 17 05:19:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	247631	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	386806	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	377009	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.288	152	157294	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	184338	49.34	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 98.68%			
35) Dibromofluoromethane	7.513	113	144311	53.80	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 107.60%			
50) Toluene-d8	10.029	98	579965	55.16	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 110.32%			
62) 4-Bromofluorobenzene	12.346	95	190084	51.42	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.84%			
Target Compounds						Qvalue
16) Acetone	3.756	43	33607	17.25	ug/l	96
18) Methyl Acetate	4.255	43	5427m	1.35	ug/l	
20) Methylene Chloride	4.469	84	10583	3.00	ug/l	94
43) Isopropyl Acetate	8.098	43	28750	3.89	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066636.D
Acq On : 16 Apr 2021 19:18
Operator : JC/MD
Sample : PB135768ZHE#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135768ZHE#06

Manual Integrations
APPROVED

MMDadoda
4/19/2021 2:59:27 PM

Quant Time: Apr 17 05:19:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

