

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066637.D
 Acq On : 16 Apr 2021 19:41
 Operator : JC/MD
 Sample : PB135768ZHE#07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135768ZHE#07

Quant Time: Apr 17 05:19:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	245088	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	390442	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	381473	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	160236	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	187351	50.67	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.34%			
35) Dibromofluoromethane	7.513	113	145535	53.75	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 107.50%			
50) Toluene-d8	10.029	98	583508	54.98	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 109.96%			
62) 4-Bromofluorobenzene	12.346	95	192020	51.46	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.92%			
Target Compounds					Qvalue	
16) Acetone	3.756	43	34516	18.24	ug/l	96
18) Methyl Acetate	4.257	43	5761	1.45	ug/l #	82
20) Methylene Chloride	4.474	84	10488	3.00	ug/l	86
43) Isopropyl Acetate	8.095	43	28938	3.88	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066637.D
Acq On : 16 Apr 2021 19:41
Operator : JC/MD
Sample : PB135768ZHE#07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135768ZHE#07

Quant Time: Apr 17 05:19:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

