

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066638.D
 Acq On : 16 Apr 2021 20:05
 Operator : JC/MD
 Sample : PB135768ZHE#08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135768ZHE#08

Manual Integrations
 APPROVED

MMDadoda
 4/19/2021 2:59:29 PM

Quant Time: Apr 17 05:20:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	255248	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	403842	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	397167	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.288	152	163475	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	193387	50.22	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.44%			
35) Dibromofluoromethane	7.513	113	149894	53.52	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 107.04%			
50) Toluene-d8	10.029	98	607834	55.38	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 110.76%			
62) 4-Bromofluorobenzene	12.346	95	199695	51.74	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 103.48%			
Target Compounds					Qvalue	
16) Acetone	3.753	43	34021	16.79	ug/l	99
18) Methyl Acetate	4.260	43	6459m	1.56	ug/l	
20) Methylene Chloride	4.467	84	11073	3.04	ug/l	97
43) Isopropyl Acetate	8.095	43	28791	3.73	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066638.D
Acq On : 16 Apr 2021 20:05
Operator : JC/MD
Sample : PB135768ZHE#08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135768ZHE#08

Manual Integrations
APPROVED

MMDadoda
4/19/2021 2:59:29 PM

Quant Time: Apr 17 05:20:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

