

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066639.D
 Acq On : 16 Apr 2021 20:29
 Operator : JC/MD
 Sample : PB135768ZHE#09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135768ZHE#09

Quant Time: Apr 17 05:20:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	261834	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	411249	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	392593	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	154073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	191109	48.38	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.76%
35) Dibromofluoromethane	7.511	113	151581	53.15	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.30%
50) Toluene-d8	10.029	98	610612	54.63	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.26%
62) 4-Bromofluorobenzene	12.346	95	194152	49.40	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.80%
Target Compounds						
					Qvalue	
16) Acetone	3.756	43	34050	16.17	ug/l	95
18) Methyl Acetate	4.260	43	6408	1.51	ug/l #	81
20) Methylene Chloride	4.477	84	11573	3.10	ug/l	90
43) Isopropyl Acetate	8.095	43	28307	3.61	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066639.D
Acq On : 16 Apr 2021 20:29
Operator : JC/MD
Sample : PB135768ZHE#09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135768ZHE#09

Quant Time: Apr 17 05:20:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

