

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066642.D
 Acq On : 16 Apr 2021 21:39
 Operator : JC/MD
 Sample : M2050-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-9

Quant Time: Apr 17 05:21:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	246410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	395039	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	381737	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	151747	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	187258	50.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.74%	
35) Dibromofluoromethane	7.513	113	146508	53.48	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.96%	
50) Toluene-d8	10.029	98	583087	54.30	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 108.60%	
62) 4-Bromofluorobenzene	12.346	95	186486	49.39	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.78%	
Target Compounds						
						Qvalue
16) Acetone	3.756	43	33833	17.56	ug/l	96
18) Methyl Acetate	4.263	43	6303	1.57	ug/l #	84
20) Methylene Chloride	4.472	84	10550	3.00	ug/l	92
43) Isopropyl Acetate	8.092	43	28447	3.77	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066642.D
Acq On : 16 Apr 2021 21:39
Operator : JC/MD
Sample : M2050-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-9

Quant Time: Apr 17 05:21:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

