

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066643.D
 Acq On : 16 Apr 2021 22:03
 Operator : JC/MD
 Sample : M2050-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Quant Time: Apr 17 05:21:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	253248	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	404034	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	394263	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.288	152	160609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	191431	50.11	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.22%	
35) Dibromofluoromethane	7.513	113	147745	52.73	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.46%	
50) Toluene-d8	10.029	98	591187	53.83	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.66%	
62) 4-Bromofluorobenzene	12.346	95	195074	50.52	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.04%	
Target Compounds						
						Qvalue
16) Acetone	3.756	43	34707	17.51	ug/l	95
18) Methyl Acetate	4.263	43	6710	1.63	ug/l #	54
20) Methylene Chloride	4.475	84	9592	2.66	ug/l	91
43) Isopropyl Acetate	8.095	43	27663	3.59	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066643.D
Acq On : 16 Apr 2021 22:03
Operator : JC/MD
Sample : M2050-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Apr 17 05:21:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

