

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066644.D
 Acq On : 16 Apr 2021 22:27
 Operator : JC/MD
 Sample : M2050-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Apr 17 05:21:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	248725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	399181	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	393621	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	174064	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	189000	50.37	ug/l	0.00
Spiked Amount 50.000	Range 61	- 141	Recovery	= 100.74%		
35) Dibromofluoromethane	7.510	113	146854	53.05	ug/l	0.00
Spiked Amount 50.000	Range 69	- 133	Recovery	= 106.10%		
50) Toluene-d8	10.029	98	584122	53.84	ug/l	0.00
Spiked Amount 50.000	Range 65	- 126	Recovery	= 107.68%		
62) 4-Bromofluorobenzene	12.346	95	197172	51.68	ug/l	0.00
Spiked Amount 50.000	Range 58	- 135	Recovery	= 103.36%		
Target Compounds						
						Qvalue
16) Acetone	3.756	43	34002	17.44	ug/l	97
18) Methyl Acetate	4.257	43	6659	1.65	ug/l #	77
20) Methylene Chloride	4.469	84	10100	2.85	ug/l	91
43) Isopropyl Acetate	8.098	43	28165	3.70	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066644.D
Acq On : 16 Apr 2021 22:27
Operator : JC/MD
Sample : M2050-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Apr 17 05:21:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

