

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
 Data File : VN081748.D
 Acq On : 16 Apr 2024 10:41
 Operator : JC\MD
 Sample : VN0416WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0416WBL01

Quant Time: Apr 17 03:20:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	262002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	456255	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	415108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	177504	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	181797	49.310	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.620%
35) Dibromofluoromethane	8.165	113	149217	49.527	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.060%
50) Toluene-d8	10.571	98	552856	49.810	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.620%
62) 4-Bromofluorobenzene	12.853	95	188797	49.776	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	99.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
Data File : VN081748.D
Acq On : 16 Apr 2024 10:41
Operator : JC\MD
Sample : VN0416WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0416WBL01

Quant Time: Apr 17 03:20:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

