

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
 Data File : VN081752.D
 Acq On : 16 Apr 2024 12:27
 Operator : JC\MD
 Sample : P2073-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RSB-01A-17-041024

Quant Time: Apr 17 03:22:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	249393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	446187	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	392519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	162972	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	171024	48.734	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.460%
35) Dibromofluoromethane	8.165	113	146938	49.871	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.740%
50) Toluene-d8	10.570	98	525836	48.445	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.880%
62) 4-Bromofluorobenzene	12.853	95	175284	47.256	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	94.520%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	8729	8.937	ug/l #	87
30) Chloroform	7.971	83	7734	1.267	ug/l	90
52) Toluene	10.629	92	76948	9.473	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
Data File : VN081752.D
Acq On : 16 Apr 2024 12:27
Operator : JC\MD
Sample : P2073-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RSB-01A-17-041024

Quant Time: Apr 17 03:22:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

