

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
 Data File : VN081756.D
 Acq On : 16 Apr 2024 14:03
 Operator : JC\MD
 Sample : P2070-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Apr 17 03:24:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	241695	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	460848	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	460297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	188062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	178025	52.344	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.680%	
35) Dibromofluoromethane	8.165	113	147005	48.306	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.620%	
50) Toluene-d8	10.571	98	594346	53.015	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.020%	
62) 4-Bromofluorobenzene	12.853	95	200264	52.273	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 104.540%	
Target Compounds						
						Qvalue
16) Acetone	4.424	43	38322	40.484	ug/l	98
37) Ethyl Acetate	7.571	43	4114	1.181	ug/l #	82
43) Isopropyl Acetate	8.688	43	289029	44.466	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
Data File : VN081756.D
Acq On : 16 Apr 2024 14:03
Operator : JC\MD
Sample : P2070-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Apr 17 03:24:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

