

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
 Data File : VN081758.D
 Acq On : 16 Apr 2024 14:51
 Operator : JC\MD
 Sample : P2073-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RMW-01B-23-041024

Quant Time: Apr 17 03:25:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	227055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	421269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	390541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	164177	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	166664	52.163	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.320%			
35) Dibromofluoromethane	8.165	113	134539	48.363	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.720%			
50) Toluene-d8	10.570	98	516793	50.428	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.860%			
62) 4-Bromofluorobenzene	12.853	95	172479	49.251	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 98.500%			
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.377	101	733869	257.820	ug/l	99
12) 1,1-Dichloroethene	4.341	96	25977	9.629	ug/l	89
30) Chloroform	7.965	83	29288	5.269	ug/l	99
32) 1,1,1-Trichloroethane	8.171	97	191626	38.877	ug/l	98
40) Benzene	8.612	78	6624	0.549	ug/l #	81
44) Trichloroethene	9.353	130	11740279	3739.343	ug/l	93
55) 1,1,2-Trichloroethane	11.018	97	2200	0.764	ug/l #	72
64) Tetrachloroethene	11.106	164	245505	74.093	ug/l	97
65) Chlorobenzene	11.894	112	5499	0.612	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
Data File : VN081758.D
Acq On : 16 Apr 2024 14:51
Operator : JC\MD
Sample : P2073-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RMW-01B-23-041024

Quant Time: Apr 17 03:25:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

