

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
 Data File : VN081761.D
 Acq On : 16 Apr 2024 16:03
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 17 03:27:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	241235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	444448	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	389448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	161456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	171745	50.594	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.180%
35) Dibromofluoromethane	8.171	113	142067	48.406	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.820%
50) Toluene-d8	10.565	98	530376	49.054	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.100%
62) 4-Bromofluorobenzene	12.853	95	171259	46.352	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	92.700%
Target Compounds						
44) Trichloroethene	9.353	130	38391	11.590	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
Data File : VN081761.D
Acq On : 16 Apr 2024 16:03
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 17 03:27:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

