

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
 Data File : VN081764.D
 Acq On : 16 Apr 2024 17:15
 Operator : JC\MD
 Sample : P2073-03DL 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RMW-01B-23-041024DL

Quant Time: Apr 17 03:28:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	219696	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	395303	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	360519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	143129	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160804	52.015	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.040%	
35) Dibromofluoromethane	8.165	113	127793	48.956	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.920%	
50) Toluene-d8	10.571	98	481951	50.117	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.240%	
62) 4-Bromofluorobenzene	12.853	95	157888	48.046	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 96.100%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.383	101	7250	2.632	ug/l	# 88
44) Trichloroethene	9.353	130	91272	30.980	ug/l	99
64) Tetrachloroethene	11.100	164	2432	0.795	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041624\
Data File : VN081764.D
Acq On : 16 Apr 2024 17:15
Operator : JC\MD
Sample : P2073-03DL 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RMW-01B-23-041024DL

Quant Time: Apr 17 03:28:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

