

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN041625\  
 Data File : VN086288.D  
 Acq On : 16 Apr 2025 10:12  
 Operator : JC\MD  
 Sample : VN0416WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0416WBL01

Quant Time: Apr 17 05:35:58 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N041525W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Apr 16 04:19:23 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 8.224  | 168  | 215943   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 9.100  | 114  | 412690   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.865 | 117  | 374988   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.788 | 152  | 162383   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 8.577  | 65    | 153932   | 49.158   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 98.320%  |
| 35) Dibromofluoromethane    | 8.165  | 113   | 105409   | 55.034   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 110.060% |
| 50) Toluene-d8              | 10.565 | 98    | 518345   | 50.637   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 101.280% |
| 62) 4-Bromofluorobenzene    | 12.847 | 95    | 186209   | 49.873   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =    | 99.740%  |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN041625\  
Data File : VN086288.D  
Acq On : 16 Apr 2025 10:12  
Operator : JC\MD  
Sample : VN0416WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0416WBL01

Quant Time: Apr 17 05:35:58 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N041525W.M  
Quant Title : SW846 8260  
QLast Update : Wed Apr 16 04:19:23 2025  
Response via : Initial Calibration

