

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041625\
 Data File : VN086291.D
 Acq On : 16 Apr 2025 12:50
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 17 03:06:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	187831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	357022	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	320997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	129310	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133036	48.844	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.680%
35) Dibromofluoromethane	8.165	113	90977	54.905	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.820%
50) Toluene-d8	10.565	98	450698	50.893	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.780%
62) 4-Bromofluorobenzene	12.847	95	155302	48.080	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041625\
Data File : VN086291.D
Acq On : 16 Apr 2025 12:50
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 17 03:06:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

