

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041625\
 Data File : VN086293.D
 Acq On : 16 Apr 2025 13:38
 Operator : JC\MD
 Sample : Q1812-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RINSE-EB-TANK-041525

Quant Time: Apr 17 03:06:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	187101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	358779	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	327128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	145489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133897	49.351	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.700%
35) Dibromofluoromethane	8.165	113	89240	53.593	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.180%
50) Toluene-d8	10.565	98	448849	50.436	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.880%
62) 4-Bromofluorobenzene	12.847	95	163259	50.296	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.600%
Target Compounds						
					Qvalue	
30) Chloroform	7.965	83	180293	40.846	ug/l	99
47) Bromodichloromethane	9.888	83	38322	10.676	ug/l	95
60) Dibromochloromethane	11.353	129	5154	1.962	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041625\
Data File : VN086293.D
Acq On : 16 Apr 2025 13:38
Operator : JC\MD
Sample : Q1812-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RINSE-EB-TANK-041525

Quant Time: Apr 17 03:06:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

