

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041625\
 Data File : VN086299.D
 Acq On : 16 Apr 2025 16:03
 Operator : JC\MD
 Sample : Q1793-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Apr 17 03:07:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	186387	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	365839	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	339739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	148157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	136113	50.360	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.720%	
35) Dibromofluoromethane	8.165	113	94036	55.383	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.760%	
50) Toluene-d8	10.565	98	460983	50.800	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.600%	
62) 4-Bromofluorobenzene	12.847	95	168047	50.772	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.540%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041625\
Data File : VN086299.D
Acq On : 16 Apr 2025 16:03
Operator : JC\MD
Sample : Q1793-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Apr 17 03:07:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

