

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041625\
 Data File : VN086303.D
 Acq On : 16 Apr 2025 17:39
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 17 03:08:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.218	168	199271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	388564	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	369369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	169507	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	146444	50.680	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.360%
35) Dibromofluoromethane	8.165	113	100601	55.785	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.560%
50) Toluene-d8	10.565	98	495275	51.387	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.780%
62) 4-Bromofluorobenzene	12.847	95	190802	54.276	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041625\
 Data File : VN086303.D
 Acq On : 16 Apr 2025 17:39
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 17 03:08:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

