

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041719\
 Data File : VN055102.D
 Acq On : 17 Apr 2019 13:31
 Operator : JC/SP
 Sample : K2197-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Apr 18 03:03:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	538485	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	871580	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	727701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	232858	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	344433	48.12	ug/l	0.00
Spiked Amount			Recovery	=	96.24%	
35) Dibromofluoromethane	7.59	113	273838	47.01	ug/l	0.00
Spiked Amount			Recovery	=	94.02%	
50) Toluene-d8	10.09	98	1043737	50.07	ug/l	0.00
Spiked Amount			Recovery	=	100.14%	
62) 4-Bromofluorobenzene	12.40	95	296473	43.74	ug/l	0.00
Spiked Amount			Recovery	=	87.48%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Acrolein	3.62	56	251	0.518	ug/l #	78
16) Acetone	3.82	43	8423	4.336	ug/l	100
18) Methyl Acetate	4.34	43	4655	0.398	ug/l #	72
20) Methylene Chloride	4.55	84	15901	2.328	ug/l	94
29) Tetrahydrofuran	7.23	42	974	0.512	ug/l #	45
30) Chloroform	7.37	83	5166	0.464	ug/l	100
31) Cyclohexane	7.66	56	10734	Below Cal	#	1
36) 1,1-Dichloropropene	7.66	75	39665	4.828	ug/l #	49
38) Carbon Tetrachloride	7.66	117	51060	6.532	ug/l #	16
43) Isopropyl Acetate	8.17	43	14620	1.418	ug/l	97
48) Methyl methacrylate	9.18	41	7867	1.540	ug/l #	17
49) 1,4-Dioxane	9.27	88	140	2.925	ug/l #	1
51) 4-Methyl-2-Pentanone	10.02	43	8335	1.400	ug/l #	39
54) cis-1,3-Dichloropropene	9.90	75	50504	5.745	ug/l #	66
57) 1,3-Dichloropropane	10.77	76	10062	1.082	ug/l #	43
58) 2-Chloroethyl Vinyl ether	9.44	63	29	10.747	ug/l #	50
59) 2-Hexanone	10.76	43	159490	42.644	ug/l #	54
71) Bromoform	12.40	173	3392	1.098	ug/l #	1
76) 1,2,3-Trichloropropane	12.40	75	151434	36.165	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	5427	0.325	ug/l	96
81) trans-1,4-Dichloro-2-buten	12.40	75	151434	144.617	ug/l #	2
84) 1,2,4-Trimethylbenzene	13.04	105	8104	0.513	ug/l	95
90) Hexachloroethane	13.89	117	620	0.248	ug/l #	6
93) 1,2,4-Trichlorobenzene	14.91	180	30	2.685	ug/l #	1
95) Naphthalene	15.13	128	4301	3.563	ug/l #	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041719\
Data File : VN055102.D
Acq On : 17 Apr 2019 13:31
Operator : JC/SP
Sample : K2197-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Apr 18 03:03:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.00 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:168 Resp: 538485

Ion Ratio Lower Upper

168 100

99 55.3 41.9 62.9



Abundance Ion 167.90 (167.60 to 168.60): VN055102.D (-1739) (-)

Ion 98.90 (98.60 to 99.60): VN055102.D (-1739) (-)

7.66

250000

200000



150000

100000

50000

0

Time-->



#13

Acrolein

Concen: 0.518 ug/l

RT: 3.62 min Scan# 573

Delta R.T. 0.01 min

Lab File: VN055102.D


Acq: 17 Apr 2019 13:31

Tgt Ion: 56 Resp: 251

Ion Ratio Lower Upper

56 100

55 55.8 59.5 89.3#



Abundance Ion 56.00 (55.70 to 56.70): VN055102.D (-478) (-)

Ion 55.00 (54.70 to 55.70): VN055102.D (-478) (-)

3.62

300

250

200


150


100

50

0

Time-->









#16
Acetone
Concen: 4.336 ug/l
RT: 3.82 min Scan# 636
Delta R.T. -0.00 min
Lab File: VN055102.D
Acq: 17 Apr 2019 13:31




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 8423
Ion Ratio Lower Upper
43 100
58 26.0 20.9 31.3



#18
Methyl Acetate
Concen: 0.398 ug/l
RT: 4.34 min Scan# 799
Delta R.T. 0.01 min
Lab File: VN055102.D
Acq: 17 Apr 2019 13:31

Tgt Ion: 43 Resp: 4655
Ion Ratio Lower Upper
43 100
74 7.2 16.1 24.1#





#20

Methvlene Chloride

Concen: 2.328 ug/l

RT: 4.55 min Scan# 863

Delta R.T. 0.01 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31

Instrument :

MSVOA_N

ClientSampled :

Tot Ion: 84 Resp: 15901


Ion Ratio Lower Upper

84 100

49 130.3 109.4 164.2

51 36.6 32.2 48.4

86 54.8 49.3 73.9






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#29

Tetrahydrofuran

Concen: 0.512 ug/l

RT: 7.23 min Scan# 1698

Delta R.T. 0.01 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31


Tgt Ion: 42 Resp: 974

Ion Ratio Lower Upper

42 100

72 6.7 29.4 44.0#



71 0.0 27.8 41.6#




Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0









#30
Chloroform
Concen: 0.464 ug/l
RT: 7.37 min Scan# 1740
Delta R.T. -0.00 min
Lab File: VN055102.D
Acq: 17 Apr 2019 13:31




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 83 Resp: 5166
Ion Ratio Lower Upper
83 100
85 65.0 51.8 77.6



#33
1,2-Dichloroethane-d4
Concen: 48.120 ug/l
RT: 8.03 min Scan# 1945
Delta R.T. 0.00 min
Lab File: VN055102.D
Acq: 17 Apr 2019 13:31

Tgt Ion: 65 Resp: 344433
Ion Ratio Lower Upper
65 100
67 52.8 0.0 106.4









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN055102.D
 Acq: 17 Apr 2019 13:31

Instrument :
 MSVOA_N
 ClientSampled :

Tot Ion	Ratio	Lower	Upper
114	100		
63	21.6	0.0	45.2
88	15.8	0.0	31.8




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 47.012 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. -0.00 min
 Lab File: VN055102.D
 Acq: 17 Apr 2019 13:31

Tgt Ion	Ratio	Lower	Upper
113	100		
111	104.5	82.1	123.1
192	21.1	16.5	24.7



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 4.828 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.13 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 75 Resp: 39665

Ion Ratio Lower Upper

75 100

110 8.1 17.5 52.5#




77 0.0 24.9 37.3#



Abundance Ion 74.90 (74.60 to 75.60): VN055102.D (-1778) (-)

Ion 109.90 (109.60 to 110.60): VN055102.D (-1778) (-)

Ion 76.90 (76.60 to 77.60): VN055102.D (-1778) (-)



#38

Carbon Tetrachloride

Concen: 6.532 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.11 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31


Tgt Ion: 117 Resp: 51060

Ion Ratio Lower Upper

117 100

119 5.3 76.3 114.5#



121 0.0 23.9 35.9#




Abundance Ion 116.80 (116.50 to 117.50): VN055102.D (-1772) (-)

Ion 118.80 (118.50 to 119.50): VN055102.D (-1772) (-)

Ion 120.80 (120.50 to 121.50): VN055102.D (-1772) (-)





#43

Isopropyl Acetate

Concen: 1.418 ug/l

RT: 8.17 min Scan# 1990

Delta R.T. 0.00 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 43 Resp: 14620

Ion Ratio Lower Upper

43 100

61 16.2 14.3 21.5




87 10.9 8.6 13.0



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#48

Methyl methacrylate

Concen: 1.540 ug/l

RT: 9.18 min Scan# 2303

Delta R.T. -0.02 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31


Tgt Ion: 41 Resp: 7867

Ion Ratio Lower Upper

41 100

69 0.0 60.5 90.7#



39 0.0 43.9 65.9#




Abundance Ion 41.00 (40.70 to 41.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 39.00 (38.70 to 39.70): VN0





#49

1,4-Dioxane

Concen: 2.925 ug/l

RT: 9.27 min Scan# 2333

Delta R.T. 0.07 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 88 Resp: 140

Ion Ratio Lower Upper




88 100

43 923852.1 31.8 47.8#

58 4205.0 61.6 92.4#



Abundance Ion 88.00 (87.70 to 88.70): VN055102.D (-2218) (-)



#50

Toluene-d8

Concen: 50.065 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN055102.D


Acq: 17 Apr 2019 13:31

Tgt Ion: 98 Resp: 1043737



Ion Ratio Lower Upper


98 100

100 63.5 51.1 76.7



Abundance Ion 98.00 (97.70 to 98.70): VN055102.D (-2493) (-)





#51

4-Methyl-2-Pentanone

Concen: 1.400 ug/l

RT: 10.02 min Scan# 2566

Delta R.T. 0.04 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 8335

Ion Ratio Lower Upper

43 100

58 0.0 29.0 43.4#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0

10.02

6000

5000

4000



3000

2000

1000

0

Time-->



#54

cis-1,3-Dichloropropene

Concen: 5.745 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. 0.07 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31


Tgt Ion: 75 Resp: 50504

Ion Ratio Lower Upper

75 100

77 0.7 25.8 38.6#

39 39.5 43.8 65.8#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0

9.90


30000


20000

10000

0

Time-->









#57
 1,3-Dichloropropane
 Concen: 1.082 ug/l
 RT: 10.77 min Scan# 2797
 Delta R.T. 0.06 min
 Lab File: VN055102.D
 Acq: 17 Apr 2019 13:31




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 76 Resp: 10062
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.6 38.4#



#58
 2-Chloroethyl Vinyl ether
 Concen: 10.747 ug/l
 RT: 9.44 min Scan# 2385
 Delta R.T. -0.25 min
 Lab File: VN055102.D
 Acq: 17 Apr 2019 13:31

Tgt Ion: 63 Resp: 29
 Ion Ratio Lower Upper
 63 100
 106 0.0 20.0 30.0#









#59
2-Hexanone
Concen: 42.644 ug/l
RT: 10.76 min Scan# 2796
Delta R.T. 0.01 min
Lab File: VN055102.D
Acq: 17 Apr 2019 13:31




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 159490
Ion Ratio Lower Upper
43 100
58 18.3 25.1 75.2#



#62
4-Bromofluorobenzene
Concen: 43.744 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.00 min
Lab File: VN055102.D
Acq: 17 Apr 2019 13:31

Tgt Ion: 95 Resp: 296473
Ion Ratio Lower Upper
95 100
174 86.2 0.0 174.4
176 82.5 0.0 166.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN055102.D
Acq: 17 Apr 2019 13:31

Instrument :
MSVOA_N
ClientSampled :

Tot Ion:117 Resp: 727701
Ion Ratio Lower Upper
117 100
82 55.2 44.6 66.8
119 32.7 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 1.098 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.27 min
Lab File: VN055102.D
Acq: 17 Apr 2019 13:31

Tgt Ion:173 Resp: 3392
Ion Ratio Lower Upper
173 100
175 508.1 24.3 73.0#
254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. 0.00 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31

Instrument :

MSVOA_N

ClientSampled :


Tot Ion:152 Resp: 232858

Ion Ratio Lower Upper

152 100

115 56.8 28.5 85.6




150 158.4 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 36.165 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN055102.D


Acq: 17 Apr 2019 13:31

Tgt Ion: 75 Resp: 151434

Ion Ratio Lower Upper



75 100


77 1.3 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): V

Ion 77.00 (76.70 to 77.70): V









#80
 1,3,5-Trimethylbenzene
 Concen: 0.325 ug/l
 RT: 12.73 min Scan# 3409
 Delta R.T. 0.00 min
 Lab File: VN055102.D
 Acq: 17 Apr 2019 13:31




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion:105 Resp: 5427
 Ion Ratio Lower Upper
 105 100
 120 46.0 24.3 72.8



#81
 trans-1,4-Dichloro-2-butene
 Concen: 144.617 ug/l
 RT: 12.40 min Scan# 3305
 Delta R.T. 0.10 min
 Lab File: VN055102.D
 Acq: 17 Apr 2019 13:31

Tgt Ion: 75 Resp: 151434
 Ion Ratio Lower Upper
 75 100
 53 0.1 96.3 144.5#
 89 0.0 38.6 57.8#





#84

1,2,4-Trimethylbenzene

Concen: 0.513 ug/l

RT: 13.04 min Scan# 3504

Delta R.T. 0.00 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:105 Resp: 8104

Ion Ratio Lower Upper




105 100

120 41.7 22.4 67.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#90

Hexachloroethane

Concen: 0.248 ug/l

RT: 13.89 min Scan# 3767

Delta R.T. 0.01 min

Lab File: VN055102.D


Acq: 17 Apr 2019 13:31

Tgt Ion:117 Resp: 620

Ion Ratio Lower Upper



117 100


201 0.0 43.4 130.2#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V





#93

1,2,4-Trichlorobenzene

Concen: 2.685 ug/l

RT: 14.91 min Scan# 4086

Delta R.T. 0.00 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31


Instrument :

MSVOA_N

ClientSampled :

Tot Ion:180 Resp: 30




Ion	Ratio	Lower	Upper
180	100		
182	0.0	47.5	142.6#
145	640.0	14.2	42.8#



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.80 (144.50 to 145.50): V



#95

Naphthalene

Concen: 3.563 ug/l

RT: 15.13 min Scan# 4155


Delta R.T. 0.00 min

Lab File: VN055102.D

Acq: 17 Apr 2019 13:31

Tgt Ion:128 Resp: 4301

Ion	Ratio	Lower	Upper
128	100		
127	7.5	10.1	15.1#
129	1.3	8.8	13.2#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

