

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041720\
 Data File : VN061053.D
 Acq On : 17 Apr 2020 16:08
 Operator : JC/MD
 Sample : L2337-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB180-EB-20200416

Quant Time: Apr 20 01:32:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	149863	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	253541	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	236820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	101519	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	101201	52.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.56%	
35) Dibromofluoromethane	7.56	113	76273	51.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.02%	
50) Toluene-d8	10.08	98	315872	56.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.92%	
62) 4-Bromofluorobenzene	12.39	95	116219	53.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.00%	
Target Compounds						
					Qvalue	
29) Tetrahydrofuran	7.18	42	246	0.280	ug/l #	36
31) Cyclohexane	7.64	56	3185	0.892	ug/l #	1
36) 1,1-Dichloropropene	7.64	75	11827	4.520	ug/l #	49
38) Carbon Tetrachloride	7.64	117	14497	6.657	ug/l #	15
51) 4-Methyl-2-Pentanone	10.07	43	1484	0.557	ug/l #	1
76) 1,2,3-Trichloropropane	12.39	75	59559	26.672	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.39	75	59559	65.789	ug/l #	12


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041720\
Data File : VN061053.D
Acq On : 17 Apr 2020 16:08
Operator : JC/MD
Sample : L2337-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleID :
BP-VPB180-EB-20200416

Quant Time: Apr 20 01:32:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1825

Delta R.T. 0.00 min

Lab File: VN061053.D

Acq: 17 Apr 2020 16:08

Instrument :

MSVOA_N

ClientSampleId :


BP-VPB180-EB-20200416

Tot Ion:168 Resp: 149863

Ion Ratio Lower Upper

168 100

99 57.6 47.4 71.2



Abundance Ion 167.90 (167.60 to 168.60): VN061006.D (-1806) (-)

Ion 98.90 (98.60 to 99.60): VN061006.D (-1806) (-)

7.64

60000



40000

20000

0

7.60 7.70

Time-->



#29

Tetrahydrofuran

Concen: 0.280 ug/l

RT: 7.18 min Scan# 1680

Delta R.T. 0.00 min

Lab File: VN061053.D

Acq: 17 Apr 2020 16:08


Tgt Ion: 42 Resp: 246

Ion Ratio Lower Upper

42 100

72 0.0 32.1 48.1#

71 0.0 30.5 45.7#



Abundance Ion 42.00 (41.70 to 42.70): VN061006.D (-1650) (-)

Ion 72.00 (71.70 to 72.70): VN061006.D (-1650) (-)

Ion 71.00 (70.70 to 71.70): VN061006.D (-1650) (-)

7.18

300


200


100

0

7.15 7.16 7.17 7.18

Time-->









#31
Cyclohexane
Concen: 0.892 ug/l
RT: 7.64 min Scan# 1824
Delta R.T. 0.01 min
Lab File: VN061053.D
Acq: 17 Apr 2020 16:08

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB180-EB-20200416

Tot Ion: 56 Resp: 3185
Ion Ratio Lower Upper
56 100
69 195.0 25.6 38.4#
84 141.9 63.4 95.2#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 52.780 ug/l
RT: 8.00 min Scan# 1937
Delta R.T. 0.00 min
Lab File: VN061053.D
Acq: 17 Apr 2020 16:08

Tgt Ion: 65 Resp: 101201
Ion Ratio Lower Upper
65 100
67 53.5 0.0 107.0



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.56 min Scan# 2112
 Delta R.T. 0.00 min
 Lab File: VN061053.D
 Acq: 17 Apr 2020 16:08




Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB180-EB-20200416


Tot Ion	Ratio	Lower	Upper
114	100		
63	23.0	0.0	46.2
88	16.2	0.0	33.4



#35
 Dibromofluoromethane
 Concen: 51.513 ug/l
 RT: 7.56 min Scan# 1801
 Delta R.T. 0.00 min
 Lab File: VN061053.D
 Acq: 17 Apr 2020 16:08

Tgt Ion	Ratio	Lower	Upper
113	100		
111	102.7	83.4	125.2
192	19.6	15.5	23.3









#36
 1,1-Dichloropropene
 Concen: 4.520 ug/l
 RT: 7.64 min Scan# 1825
 Delta R.T. -0.13 min
 Lab File: VN061053.D
 Acq: 17 Apr 2020 16:08




Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB180-EB-20200416


Tot Ion	75	Resp	11827
Ion Ratio	100	Lower	Upper
75	100		
110	6.9	16.8	50.3#
77	0.0	24.7	37.1#



#38
 Carbon Tetrachloride
 Concen: 6.657 ug/l
 RT: 7.64 min Scan# 1825
 Delta R.T. -0.11 min
 Lab File: VN061053.D
 Acq: 17 Apr 2020 16:08

Tgt Ion	117	Resp	14497
Ion Ratio	100	Lower	Upper
117	100		
119	4.8	78.1	117.1#
121	0.0	24.0	36.0#





#50

Toluene-d8

Concen: 56.455 ug/l

RT: 10.08 min Scan# 2582

Delta R.T. 0.00 min

Lab File: VN061053.D

Acq: 17 Apr 2020 16:08

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB180-EB-20200416

Tot Ion: 98 Resp: 315872

Ion Ratio Lower Upper

98 100

100 64.2 51.9 77.9



#51

4-Methyl-2-Pentanone

Concen: 0.557 ug/l

RT: 10.07 min Scan# 2581

Delta R.T. 0.11 min

Lab File: VN061053.D




Acq: 17 Apr 2020 16:08


Tgt Ion: 43 Resp: 1484

Ion Ratio Lower Upper

43 100

58 177.6 30.3 45.5#





#62

4-Bromofluorobenzene

Concen: 52.999 ug/l

RT: 12.39 min Scan# 3303

Delta R.T. 0.00 min

Lab File: VN061053.D

Acq: 17 Apr 2020 16:08

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB180-EB-20200416


Tot Ion: 95 Resp: 116219

Ion Ratio Lower Upper

95 100

174 75.4 0.0 151.2




176 73.5 0.0 148.2



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): VN0

Ion 175.80 (175.50 to 176.50): VN0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.40 min Scan# 2994

Delta R.T. 0.00 min

Lab File: VN061053.D

Acq: 17 Apr 2020 16:08


Tgt Ion: 117 Resp: 236820

Ion Ratio Lower Upper

117 100

82 58.9 47.1 70.7



119 32.3 25.7 38.5




Abundance Ion 116.90 (116.60 to 117.60): VN0

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): VN0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. 0.00 min

Lab File: VN061053.D

Acq: 17 Apr 2020 16:08

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB180-EB-20200416


Tot Ion:152 Resp: 101519

Ion Ratio Lower Upper

152 100

115 61.1 30.8 92.4




150 158.4 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 26.672 ug/l

RT: 12.39 min Scan# 3303

Delta R.T. -0.15 min

Lab File: VN061053.D


Acq: 17 Apr 2020 16:08

Tgt Ion: 75 Resp: 59559

Ion Ratio Lower Upper



75 100


77 0.9 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0





#81

trans-1,4-Dichloro-2-butene

Concen: 65.789 ug/l

RT: 12.39 min Scan# 3303

Delta R.T. 0.10 min

Lab File: VN061053.D

Acq: 17 Apr 2020 16:08

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB180-EB-20200416

Tot Ion: 75 Resp: 59559

Ion Ratio Lower Upper

75 100

53 0.0 74.7 112.1#

89 0.0 34.6 52.0#

