

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041720\
 Data File : VN061048.D
 Acq On : 17 Apr 2020 14:15
 Operator : JC/MD
 Sample : L2329-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-5

Quant Time: Apr 20 03:31:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	151819	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	259706	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	240563	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	105009	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	102608	52.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.64%	
35) Dibromofluoromethane	7.56	113	78056	51.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.94%	
50) Toluene-d8	10.08	98	318129	55.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.02%	
62) 4-Bromofluorobenzene	12.40	95	117620	52.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041720\
Data File : VN061048.D
Acq On : 17 Apr 2020 14:15
Operator : JC/MD
Sample : L2329-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-5

Quant Time: Apr 20 03:31:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

