

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041720\
 Data File : VN061059.D
 Acq On : 17 Apr 2020 18:25
 Operator : JC/MD
 Sample : L2337-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB180-GW-483-485

Quant Time: Apr 20 04:08:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	138584	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	235183	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	222559	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	96846	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	95087	53.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.26%	
35) Dibromofluoromethane	7.56	113	71593	52.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.26%	
50) Toluene-d8	10.08	98	294543	56.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.50%	
62) 4-Bromofluorobenzene	12.40	95	107298	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
Target Compounds						
16) Acetone	3.79	43	3420	4.382	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041720\
Data File : VN061059.D
Acq On : 17 Apr 2020 18:25
Operator : JC/MD
Sample : L2337-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB180-GW-483-485

Quant Time: Apr 20 04:08:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

