

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041720\
 Data File : VN061068.D
 Acq On : 17 Apr 2020 22:34
 Operator : JC/MD
 Sample : VN0417WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0417WBL02

Quant Time: Apr 20 06:06:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	143865	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	244850	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	227291	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	98934	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	99133	53.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.72%	
35) Dibromofluoromethane	7.56	113	73975	51.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.46%	
50) Toluene-d8	10.08	98	303804	56.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.46%	
62) 4-Bromofluorobenzene	12.40	95	110627	52.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.48%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041720\
Data File : VN061068.D
Acq On : 17 Apr 2020 22:34
Operator : JC/MD
Sample : VN0417WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0417WBL02

Quant Time: Apr 20 06:06:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

