

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041720\
 Data File : VN061071.D
 Acq On : 17 Apr 2020 23:42
 Operator : JC/MD
 Sample : L2334-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Z3-TB6-200415

Quant Time: Apr 20 06:21:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	143151	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	245309	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	227463	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	97352	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	97761	53.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.76%	
35) Dibromofluoromethane	7.56	113	73243	51.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.26%	
50) Toluene-d8	10.08	98	303636	56.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.18%	
62) 4-Bromofluorobenzene	12.39	95	109821	51.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.52%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041720\
Data File : VN061071.D
Acq On : 17 Apr 2020 23:42
Operator : JC/MD
Sample : L2334-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Z3-TB6-200415

Quant Time: Apr 20 06:21:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

