

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041720\
 Data File : VN061070.D
 Acq On : 17 Apr 2020 23:19
 Operator : JC/MD
 Sample : L2338-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TB01-20200414

Quant Time: Apr 20 06:18:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	141253	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	240243	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	218021	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	94048	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	96424	53.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.70%	
35) Dibromofluoromethane	7.56	113	72921	51.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.96%	
50) Toluene-d8	10.08	98	297146	56.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.10%	
62) 4-Bromofluorobenzene	12.40	95	106404	51.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.42%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041720\
Data File : VN061070.D
Acq On : 17 Apr 2020 23:19
Operator : JC/MD
Sample : L2338-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TB01-20200414

Quant Time: Apr 20 06:18:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1825

Delta R.T. 0.00 min

Lab File: VN061070.D

Acq: 17 Apr 2020 23:19

Instrument :

MSVOA_N

ClientSampleId :


BP-TB01-20200414

Tot Ion:168 Resp: 141253

Ion Ratio Lower Upper




168 100

99 57.4 47.4 71.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#33

1,2-Dichloroethane-d4

Concen: 53.354 ug/l

RT: 8.00 min Scan# 1937

Delta R.T. 0.00 min

Lab File: VN061070.D


Acq: 17 Apr 2020 23:19

Tgt Ion: 65 Resp: 96424

Ion Ratio Lower Upper



65 100


67 54.0 0.0 107.0



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.56 min Scan# 2112

Delta R.T. 0.00 min

Lab File: VN061070.D

Acq: 17 Apr 2020 23:19

Instrument :

MSVOA_N

ClientSampleId :

BP-TB01-20200414


Tot Ion:114 Resp: 240243

Ion Ratio Lower Upper

114 100

63 23.3 0.0 46.2




88 16.2 0.0 33.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 51.975 ug/l

RT: 7.56 min Scan# 1801

Delta R.T. 0.00 min

Lab File: VN061070.D

Acq: 17 Apr 2020 23:19


Tgt Ion:113 Resp: 72921

Ion Ratio Lower Upper

113 100

111 103.6 83.4 125.2



192 19.4 15.5 23.3




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 56.048 ug/l

RT: 10.08 min Scan# 2582

Delta R.T. 0.00 min

Lab File: VN061070.D

Acq: 17 Apr 2020 23:19

Instrument :

MSVOA_N

ClientSampleId :





BP-TB01-20200414

Tot Ion: 98 Resp: 297146

Ion Ratio Lower Upper

98 100

100 65.1 51.9 77.9



#62

4-Bromofluorobenzene

Concen: 51.209 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. 0.00 min

Lab File: VN061070.D

Acq: 17 Apr 2020 23:19




Tgt Ion: 95 Resp: 106404


Ion Ratio Lower Upper

95 100

174 75.5 0.0 151.2

176 73.5 0.0 148.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2994
Delta R.T. 0.00 min
Lab File: VN061070.D
Acq: 17 Apr 2020 23:19

Instrument :
MSVOA_N
ClientSampleId :
BP-TB01-20200414

Tot Ion:117 Resp: 218021
Ion Ratio Lower Upper
117 100
82 59.1 47.1 70.7
119 32.4 25.7 38.5




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3597
Delta R.T. 0.00 min
Lab File: VN061070.D
Acq: 17 Apr 2020 23:19

Tgt Ion:152 Resp: 94048
Ion Ratio Lower Upper
152 100
115 59.2 30.8 92.4
150 156.6 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

