


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN041720\
 Data File : VN061066.D
 Acq On : 17 Apr 2020 21:26
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Title : SW846 8260
 Last Update : Fri Apr 17 03:49:39 2020



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	6047	PASS
75	95	30	60	50.8	14921	PASS
95	95	100	100	100.0	29386	PASS
96	95	5	9	6.7	1977	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.7	22253	PASS
175	174	5	9	7.3	1626	PASS
176	174	95	101	96.0	21352	PASS
177	176	5	9	6.6	1413	PASS