

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041723\
 Data File : VN077351.D
 Acq On : 17 Apr 2023 11:25
 Operator : JC\MD
 Sample : VN0417MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0417MBL01

Quant Time: Apr 18 04:39:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	461592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	863112	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	731495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	240237	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	282734	44.816	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.640%
35) Dibromofluoromethane	8.172	113	262000	47.418	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.840%
50) Toluene-d8	10.572	98	1044099	46.940	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.880%
62) 4-Bromofluorobenzene	12.854	95	300341	40.598	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.200%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041723\
Data File : VN077351.D
Acq On : 17 Apr 2023 11:25
Operator : JC\MD
Sample : VN0417MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0417MBL01

Quant Time: Apr 18 04:39:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

