

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041723\
 Data File : VN077359.D
 Acq On : 17 Apr 2023 16:05
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 18 04:41:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	535193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1000730	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	883406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	310005	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	324519	44.365	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.740%
35) Dibromofluoromethane	8.172	113	300268	46.870	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.740%
50) Toluene-d8	10.571	98	1217226	47.198	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.400%
62) 4-Bromofluorobenzene	12.854	95	361741	42.174	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041723\
Data File : VN077359.D
Acq On : 17 Apr 2023 16:05
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 18 04:41:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

