

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041725\
 Data File : VN086323.D
 Acq On : 17 Apr 2025 15:42
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 18 01:32:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	213344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	413360	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	379743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	172906	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	152250	49.213	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.420%
35) Dibromofluoromethane	8.165	113	106643	55.588	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.180%
50) Toluene-d8	10.565	98	518432	50.563	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.120%
62) 4-Bromofluorobenzene	12.847	95	191313	51.156	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.320%
Target Compounds						
					Qvalue	
68) m/p-Xylenes	12.065	106	8447	1.474	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041725\
Data File : VN086323.D
Acq On : 17 Apr 2025 15:42
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 18 01:32:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

