

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041725\
 Data File : VN086324.D
 Acq On : 17 Apr 2025 16:06
 Operator : JC\MD
 Sample : Q1818-01DL 10X
 Misc : 1.14g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 3822DL

Quant Time: Apr 18 01:32:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	201843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	389459	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	358086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	162964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	145270	49.633	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.260%		
35) Dibromofluoromethane	8.165	113	101768	56.302	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	112.600%		
50) Toluene-d8	10.565	98	491426	50.871	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.740%		
62) 4-Bromofluorobenzene	12.847	95	181993	51.651	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	103.300%		
Target Compounds						Qvalue
39) Methylcyclohexane	9.600	83	4438	1.034	ug/l	90
52) Toluene	10.629	92	402199	55.725	ug/l	98
67) Ethyl Benzene	11.959	91	4941	0.343	ug/l	92
68) m/p-Xylenes	12.065	106	7213	1.335	ug/l	96
69) o-Xylene	12.394	106	1877	0.351	ug/l	77
84) 1,2,4-Trimethylbenzene	13.476	105	10058	0.906	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041725\
Data File : VN086324.D
Acq On : 17 Apr 2025 16:06
Operator : JC\MD
Sample : Q1818-01DL 10X
Misc : 1.14g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
3822DL

Quant Time: Apr 18 01:32:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

