

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN041819\
 Data File : VN055114.D
 Acq On : 18 Apr 2019 9:48
 Operator : JC/SP
 Sample : VN0418MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0418MBL01

Quant Time: Apr 19 02:03:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	517156	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	834349	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	686745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	218117	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	332849	48.42	ug/l	0.00
Spiked Amount			Recovery	=	96.84%	
35) Dibromofluoromethane	7.59	113	266173	47.74	ug/l	0.00
Spiked Amount			Recovery	=	95.48%	
50) Toluene-d8	10.09	98	991428	49.68	ug/l	0.00
Spiked Amount			Recovery	=	99.36%	
62) 4-Bromofluorobenzene	12.40	95	284764	43.89	ug/l	0.00
Spiked Amount			Recovery	=	87.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN041819\
Data File : VN055114.D
Acq On : 18 Apr 2019 9:48
Operator : JC/SP
Sample : VN0418MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0418MBL01

Quant Time: Apr 19 02:03:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

