

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041819\
 Data File : VN055116.D
 Acq On : 18 Apr 2019 10:46
 Operator : JC/SP
 Sample : PB118998TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB118998TB

Quant Time: Apr 19 02:16:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	404188	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	639139	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	545551	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	167734	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	265174	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
35) Dibromofluoromethane	7.59	113	211187	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
50) Toluene-d8	10.09	98	778882	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	12.40	95	220302	44.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.66%	
Target Compounds						
20) Methylene Chloride	4.54	84	15065	2.939	ug/l	89
43) Isopropyl Acetate	8.17	43	14759	1.953	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041819\
Data File : VN055116.D
Acq On : 18 Apr 2019 10:46
Operator : JC/SP
Sample : PB118998TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB118998TB

Quant Time: Apr 19 02:16:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

