

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041819\
 Data File : VN055118.D
 Acq On : 18 Apr 2019 11:39
 Operator : JC/SP
 Sample : K2461-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PZ-01

Quant Time: Apr 19 02:29:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	340999	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	533112	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	435650	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	122673	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	217988	48.09	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.18%	
35) Dibromofluoromethane	7.59	113	173480	48.69	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.38%	
50) Toluene-d8	10.09	98	639257	50.13	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.26%	
62) 4-Bromofluorobenzene	12.40	95	168922	40.75	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	81.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041819\
Data File : VN055118.D
Acq On : 18 Apr 2019 11:39
Operator : JC/SP
Sample : K2461-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PZ-01

Quant Time: Apr 19 02:29:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

