

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041819\
 Data File : VN055120.D
 Acq On : 18 Apr 2019 14:55
 Operator : JC/SP
 Sample : K2461-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Apr 19 02:34:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	576641	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	883784	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	739552	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	225517	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	347753	45.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.74%	
35) Dibromofluoromethane	7.59	113	285141	48.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.56%	
50) Toluene-d8	10.09	98	1065078	50.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.76%	
62) 4-Bromofluorobenzene	12.40	95	307362	44.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.44%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041819\
Data File : VN055120.D
Acq On : 18 Apr 2019 14:55
Operator : JC/SP
Sample : K2461-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Apr 19 02:34:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

