

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041819\
 Data File : VN055124.D
 Acq On : 18 Apr 2019 16:36
 Operator : JC/SP
 Sample : K2193-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HR-05-041719-C

Quant Time: Apr 19 03:02:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	553619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	873958	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	718000	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	241805	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	353526	48.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.08%	
35) Dibromofluoromethane	7.59	113	278101	47.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.22%	
50) Toluene-d8	10.09	98	1023953	48.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.96%	
62) 4-Bromofluorobenzene	12.40	95	302588	44.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.04%	
Target Compounds						
16) Acetone	3.82	43	15000	7.510	ug/l	93
20) Methylene Chloride	4.55	84	591989	84.312	ug/l	97
30) Chloroform	7.38	83	13026	1.139	ug/l	93
43) Isopropyl Acetate	8.17	43	17375	1.681	ug/l	95
80) 1,3,5-Trimethylbenzene	12.73	105	40485	2.338	ug/l	97
84) 1,2,4-Trimethylbenzene	13.04	105	67888	4.137	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041819\
Data File : VN055124.D
Acq On : 18 Apr 2019 16:36
Operator : JC/SP
Sample : K2193-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HR-05-041719-C

Quant Time: Apr 19 03:02:46 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

