

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041824\
 Data File : VN081778.D
 Acq On : 18 Apr 2024 11:19
 Operator : JC\MD
 Sample : VN0418WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0418WBL01

Quant Time: Apr 19 01:19:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	191033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	337160	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	316874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122586	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	146079	54.342	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.680%	
35) Dibromofluoromethane	8.165	113	115776	52.001	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.000%	
50) Toluene-d8	10.565	98	424991	51.815	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.640%	
62) 4-Bromofluorobenzene	12.853	95	128044	45.683	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 91.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041824\
Data File : VN081778.D
Acq On : 18 Apr 2024 11:19
Operator : JC\MD
Sample : VN0418WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0418WBL01

Quant Time: Apr 19 01:19:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

