

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041824\
 Data File : VN081782.D
 Acq On : 18 Apr 2024 13:06
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 19 01:21:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	235327	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	420879	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	410671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	196806	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	166487	50.276	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.560%	
35) Dibromofluoromethane	8.171	113	137788	49.577	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.160%	
50) Toluene-d8	10.565	98	536163	52.366	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.740%	
62) 4-Bromofluorobenzene	12.847	95	215306	61.537	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 123.080%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041824\
Data File : VN081782.D
Acq On : 18 Apr 2024 13:06
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 19 01:21:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

