

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041824\
 Data File : VN081792.D
 Acq On : 18 Apr 2024 17:06
 Operator : JC\MD
 Sample : P2121-05 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 4827

Quant Time: Apr 19 01:26:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	226172	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	395466	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	367382	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	162091	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	164878	51.806	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.620%
35) Dibromofluoromethane	8.165	113	130451	49.954	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.900%
50) Toluene-d8	10.571	98	495399	51.494	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.980%
62) 4-Bromofluorobenzene	12.853	95	172971	52.614	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	105.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041824\
Data File : VN081792.D
Acq On : 18 Apr 2024 17:06
Operator : JC\MD
Sample : P2121-05 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
4827

Quant Time: Apr 19 01:26:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

