

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041824\
 Data File : VN081799.D
 Acq On : 18 Apr 2024 19:54
 Operator : JC\MD
 Sample : P2185-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOTE-1

Quant Time: Apr 19 02:12:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	234098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	389818	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	141900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	61389	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159973	48.563	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.120%
35) Dibromofluoromethane	8.165	113	131173	50.958	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.920%
50) Toluene-d8	10.571	98	192870	20.338	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	40.680%#
62) 4-Bromofluorobenzene	12.853	95	57370	17.703	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	35.400%#
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	3575	3.899	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041824\
Data File : VN081799.D
Acq On : 18 Apr 2024 19:54
Operator : JC\MD
Sample : P2185-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOTE-1

Quant Time: Apr 19 02:12:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

