

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
 Data File : VN086341.D
 Acq On : 18 Apr 2025 14:07
 Operator : JC\MD
 Sample : PB167623ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167623ZHE#06

Quant Time: Apr 19 02:24:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	196766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	376539	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	353824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	154059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	139583	48.920	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.840%
35) Dibromofluoromethane	8.165	113	92133	52.721	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.440%
50) Toluene-d8	10.565	98	478953	51.281	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.560%
62) 4-Bromofluorobenzene	12.847	95	175618	51.552	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.100%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	51733	47.826	ug/l	99
18) Methyl Acetate	5.024	43	4032	1.177	ug/l #	86
20) Methylene Chloride	5.277	84	5816	2.205	ug/l #	91
43) Isopropyl Acetate	8.682	43	49612	5.761	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
Data File : VN086341.D
Acq On : 18 Apr 2025 14:07
Operator : JC\MD
Sample : PB167623ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167623ZHE#06

Quant Time: Apr 19 02:24:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

