

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
 Data File : VN086344.D
 Acq On : 18 Apr 2025 15:19
 Operator : JC\MD
 Sample : PB167623ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167623ZHE#09

Quant Time: Apr 19 02:26:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	197092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	376730	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	355138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	154079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	139743	48.895	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.800%
35) Dibromofluoromethane	8.165	113	90973	52.031	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.060%
50) Toluene-d8	10.565	98	476557	50.998	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.000%
62) 4-Bromofluorobenzene	12.847	95	174663	51.245	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.500%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	35472	30.896	ug/l	99
18) Methyl Acetate	5.030	43	3718	1.084	ug/l #	81
20) Methylene Chloride	5.283	84	4229	1.601	ug/l	86
43) Isopropyl Acetate	8.688	43	48565	5.595	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
Data File : VN086344.D
Acq On : 18 Apr 2025 15:19
Operator : JC\MD
Sample : PB167623ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167623ZHE#09

Quant Time: Apr 19 02:26:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

