

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
 Data File : VN086345.D
 Acq On : 18 Apr 2025 15:43
 Operator : JC\MD
 Sample : PB167623ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167623ZHE#10

Quant Time: Apr 19 02:26:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	195047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	376446	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	348966	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	154510	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	140325	49.614	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.220%
35) Dibromofluoromethane	8.159	113	91680	52.474	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.940%
50) Toluene-d8	10.565	98	479043	51.303	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.600%
62) 4-Bromofluorobenzene	12.847	95	171606	50.387	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.780%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	36063	31.900	ug/l	98
18) Methyl Acetate	5.018	43	3939	1.160	ug/l #	76
20) Methylene Chloride	5.271	84	4373	1.672	ug/l	96
43) Isopropyl Acetate	8.688	43	45409	5.114	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
Data File : VN086345.D
Acq On : 18 Apr 2025 15:43
Operator : JC\MD
Sample : PB167623ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167623ZHE#10

Quant Time: Apr 19 02:26:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

