

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
 Data File : VN086346.D
 Acq On : 18 Apr 2025 16:07
 Operator : JC\MD
 Sample : PB167623ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167623ZHE#11

Quant Time: Apr 19 02:27:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	190477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	365011	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	344723	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	151460	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	137077	49.628	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.260%
35) Dibromofluoromethane	8.165	113	89817	53.019	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.040%
50) Toluene-d8	10.565	98	464729	51.329	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.660%
62) 4-Bromofluorobenzene	12.847	95	169237	51.248	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.500%
Target Compounds						
					Qvalue	
13) Acrolein	4.194	56	56	Below Cal	#	13
16) Acetone	4.424	43	35238	31.921	ug/l	99
18) Methyl Acetate	5.035	43	4056	1.223	ug/l #	73
20) Methylene Chloride	5.271	84	4374	1.713	ug/l	97
43) Isopropyl Acetate	8.688	43	44790	5.235	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
Data File : VN086346.D
Acq On : 18 Apr 2025 16:07
Operator : JC\MD
Sample : PB167623ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167623ZHE#11

Quant Time: Apr 19 02:27:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

