

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
 Data File : VN086348.D
 Acq On : 18 Apr 2025 16:55
 Operator : JC\MD
 Sample : PB167623ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167623ZHE#13

Quant Time: Apr 19 02:28:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	188035	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	362239	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	335998	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	147603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	137538	50.442	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.880%			
35) Dibromofluoromethane	8.165	113	89471	53.219	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.440%			
50) Toluene-d8	10.565	98	460459	51.247	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.500%			
62) 4-Bromofluorobenzene	12.847	95	167736	51.182	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.360%			
Target Compounds						
16) Acetone	4.436	43	34209	31.295	ug/l	96
18) Methyl Acetate	5.030	43	4066	1.242	ug/l #	82
20) Methylene Chloride	5.277	84	4395	1.744	ug/l #	84
43) Isopropyl Acetate	8.688	43	40251	4.561	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041825\
Data File : VN086348.D
Acq On : 18 Apr 2025 16:55
Operator : JC\MD
Sample : PB167623ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167623ZHE#13

Quant Time: Apr 19 02:28:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

