

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041919\
 Data File : VN055136.D
 Acq On : 19 Apr 2019 12:40
 Operator : JC/SP
 Sample : K2479-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP

Quant Time: Apr 22 10:01:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N041119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 12 07:55:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	69660	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	357446	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	289346	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	161689	30.45	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.50%
60) 4-Bromofluorobenzene	12.40	95	100291	22.26	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	74.20%
63) Toluene-d8	10.09	98	427399	31.07	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	103.57%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041919\
Data File : VN055136.D
Acq On : 19 Apr 2019 12:40
Operator : JC/SP
Sample : K2479-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP

Quant Time: Apr 22 10:01:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N041119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Apr 12 07:55:25 2019
Response via : Initial Calibration

