

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041919\
 Data File : VN055149.D
 Acq On : 19 Apr 2019 19:56
 Operator : JC/SP
 Sample : K2201-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-01-041819-E

Quant Time: Apr 20 04:53:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	534061	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	836405	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	671536	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	198019	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	338166	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
35) Dibromofluoromethane	7.59	113	264099	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
50) Toluene-d8	10.09	98	977709	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	12.40	95	270933	41.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.32%	
Target Compounds						
16) Acetone	3.82	43	10196	5.292	ug/l	97
20) Methylene Chloride	4.55	84	16451	2.429	ug/l	87
43) Isopropyl Acetate	8.17	43	15915	1.609	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041919\
Data File : VN055149.D
Acq On : 19 Apr 2019 19:56
Operator : JC/SP
Sample : K2201-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-01-041819-E

Quant Time: Apr 20 04:53:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

