

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041919\
 Data File : VN055150.D
 Acq On : 19 Apr 2019 20:21
 Operator : JC/SP
 Sample : K2201-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-01-041819-D

Quant Time: Apr 20 04:56:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	535408	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	846038	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	689393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	212486	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	345044	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
35) Dibromofluoromethane	7.59	113	276361	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
50) Toluene-d8	10.09	98	998512	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
62) 4-Bromofluorobenzene	12.40	95	282002	42.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.72%	
Target Compounds						
16) Acetone	3.83	43	11010	5.700	ug/l	94
20) Methylene Chloride	4.55	84	1601287	235.816	ug/l	97
43) Isopropyl Acetate	8.17	43	15585	1.558	ug/l #	95
95) Naphthalene	15.14	128	5955	3.813	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041919\
Data File : VN055150.D
Acq On : 19 Apr 2019 20:21
Operator : JC/SP
Sample : K2201-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-01-041819-D

Quant Time: Apr 20 04:56:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

