

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041919\
 Data File : VN055155.D
 Acq On : 19 Apr 2019 22:29
 Operator : JC/SP
 Sample : K2475-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-TRACK-78-79

Quant Time: Apr 20 05:10:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	539201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	858106	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	691940	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	203724	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	349046	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
35) Dibromofluoromethane	7.59	113	274777	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
50) Toluene-d8	10.09	98	1011993	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
62) 4-Bromofluorobenzene	12.40	95	284898	42.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.40%	
Target Compounds						
16) Acetone	3.82	43	13836	7.112	ug/l #	90
20) Methylene Chloride	4.55	84	1111425	162.524	ug/l	98
43) Isopropyl Acetate	8.17	43	17228	1.698	ug/l #	96
95) Naphthalene	15.13	128	7820	4.084	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041919\
Data File : VN055155.D
Acq On : 19 Apr 2019 22:29
Operator : JC/SP
Sample : K2475-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-TRACK-78-79

Quant Time: Apr 20 05:10:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

