

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041919\
 Data File : VN055157.D
 Acq On : 19 Apr 2019 23:20
 Operator : JC/SP
 Sample : K2475-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-TRACK-70-71

Quant Time: Apr 20 05:16:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	540776	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	849626	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	689173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	207042	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	343722	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
35) Dibromofluoromethane	7.59	113	269240	47.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.84%	
50) Toluene-d8	10.09	98	992912	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	12.40	95	281804	42.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.30%	
Target Compounds						
8) Diethyl Ether	3.41	74	8882	3.302	ug/l	82
16) Acetone	3.83	43	17598	9.020	ug/l #	90
20) Methylene Chloride	4.56	84	127337	18.566	ug/l	97
43) Isopropyl Acetate	8.18	43	17790	1.771	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041919\
Data File : VN055157.D
Acq On : 19 Apr 2019 23:20
Operator : JC/SP
Sample : K2475-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-TRACK-70-71

Quant Time: Apr 20 05:16:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

